

Verification of Compliance

Product Name : **Medical Type Switching Power Supply**
Brand Name : **MEAN WELL**
Model Number : **GSM40Ax(x=05, 07, 09, 12, 15, 18, 24, 28, 48),
GSM60Ax(x=05, 07, 09, 12, 15, 18, 24, 48), GSM40A48-CIS**
Applicant : **MEAN WELL ENTERPRISES CO., LTD.**
Address : **No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan
(R.O.C.)**
Report Number : **S8C-M020-1302-120**
Issue Date : **September 15, 2023**

Applicable Standards :

Emission :

EN 60601-1-2:2015/A1:2021:
Home Healthcare Environment
EN 55011:2016/A2:2021 Group 1, Class B
EN 55032: 2015/A11:2020 Class B
EN IEC 61000-3-2:2019
EN 61000-3-3:2013/A1:2019
EN IEC 61204-3:2018 Class B

Immunity :

EN 60601-1-2:2015/A1:2021: Home Healthcare
Environment
EN IEC 61204-3:2018: Residential commercial and light
Industrial environment
EN 61000-4-2:2009
EN IEC 61000-4-3:2020
EN 61000-4-4:2012
EN 61000-4-5:2014/A1:2017
EN 61000-4-6:2014
EN 61000-4-8:2010
EN IEC 61000-4-11:2020



Based on the EMC Directive 2014/30/EU, MDR (Regulation (EU) 2017/745) and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



Central Research Technology Co.

11, Lane 41, Fushuen St., Jungshan Chiu,
Taipei 104, Taiwan

Tel : 886-2-25872719

Fax : 886-2-25850364

(Jim Chien/Authorized Signatory)

Date: September 15, 2023

TAF 0905

FCC CAB Code TW1104, TW0019

NVLAP Lab Code 200575-0

ISED CAB Code TW0905

VCCI Accep. No. R-11527, C-11609, T-11441, G-10010, C-20010

T-20009, G-10614