


VERIFICATION OF COMPLIANCE

Verification No.: KSEM230800190601ATC
Applicant: MEAN WELL Enterprises Co., Ltd.
Address of Applicant: No.28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891, Taiwan
Product Description: AC/DC MEDICAL ADAPTOR
Model No.: GSM90Ax (x=12,15,19,24,48)
Sufficient samples of the product have been tested and found to be in conformity with
Test Standards: EN 55011:2016+A2:2021
EN 60601-1-2:2015+A1:2021
EN IEC 61204-3:2018
EN IEC 61000-3-2: 2019+A1:2021
EN 61000-3-3: 2013+ A1:2019+A2:2021
As shown in the
Test Report Number(s): KSEM230800190601

This verification of EMC Compliance has been granted to the applicant based on the results of the tests, performed by laboratory of Compliance Certification Services (Kunshan) Inc. on the sample of the above-mentioned product in accordance with the provisions of the relevant specific standards under Directive 2014/30/EU. The CE mark as shown below can be used, under the responsibility of the manufacturer, after completion of an EU Declaration of Conformity and compliance with all relevant EU Directives.



Eric Lin
Laboratory Manager



Date: 2023-08-28

Copyright of this verification is owned by Compliance Certification Services (Kunshan) Inc. and may not be reproduced other than in full and with the prior approval of the General Manager. This verification is subjected to the governance of the General Conditions of Services which can be accessed at <https://www.sgs.com/en/terms-and-conditions>