

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching power supply

Name and address of the applicant

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

Name and address of the manufacturer

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

Name and address of the factory

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

When more than one factory

 Additional information on page 2

Ratings and principal characteristics

Product data

Ratings: Input: 100-240 Vac, 277 Vac, 50/60 Hz;
Output: See General product information of report for details.

Product data – type NSP-100-xzzzz

Rated input current: 1,2-0,5 A (when input 100-240Vac);
0,45 A (when input 277Vac)

Product data – type NSP-75-xzzzz

Rated input current: 0,9-0,4 A (when input 100-240Vac);
0,35 A (when input 277Vac)

Trademark / Brand (if any)



MEAN WELL

Customer's Testing Facility (CTF) Stage
used

NSP-100-xzzzz and NSP-75-xzzzz

Model / Type Ref.

 Additional information on page 2

Additional information

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and
IEC 60601-1:2005/AMD2:2020

National differences:

CA, CH, CN, EU Group Differences, EU Special National Conditions, JP,
KR, USAs shown in the Test Report Ref. No.
which forms part of this Certificate

4941907.50

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.
Meander 1051
6825 MJ Arnhem
Netherlands

Date: 2025-12-17



Signature: Rosa Zhou



IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME[Additional factory](#)

MEAN WELL (GUANGZHOU) Electronics Co., Ltd.
No.11 Jingu South Road, Huadu District, 510890 Guangzhou Guangdong, China

YONGDEN TECHNOLOGY CORPORATION
345 MacArthur HighWay, Tabang, 3015 Guiguinto, Bulacan, Philippines

SuZhou MEAN WELL Technology Co., Ltd.
No.269, Changping Road, Huangdai Town, Xiangcheng District, 215152 Suzhou Jiangsu, China

MEAN WELL INDIA ELECTRONICS PRIVATE LIMITED
9C, Peenya industrial area, Chokkasandra, 2ND PHASE, PEENYA, 560058 Bengaluru (BANGALORE) Urban, Karnataka, India

[Additional information](#)

Model/Type reference:

NSP-100-xzzzz

(x=5, 7.5, 12, 15, 24, 27, 36, 48, 60; z=may be blank. -, 0~9, A~Z, a~z for market purpose)

NSP-75-xzzzz

(x=5, 12, 15, 24, 27, 36, 48, 60; z=may be blank. -, 0~9, A~Z, a~z for market purpose)

Only the requirements of IEC 60601-1 as documented in the accompanying test report, have been addressed.

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.
Meander 1051
6825 MJ Arnhem
Netherlands

Date: 2025-12-17



Signature: Rosa Zhou

