



ATTESTATION of conformity with European Directives

Attestation Number: 2588AB070201001
Product: switching power supplies
Brand Name: MEAN WELL
Model: NSP-75-xzzzz (x=5, 12, 15, 24, 27, 36, 48, 60; z=may be blank, -, 0~9, A~Z, a~z for market purpose)
Additional Models: NSP-100-xzzzz (x=5, 7.5, 12, 15, 24, 27, 36, 48, 60; z=may be blank, -, 0~9, A~Z, a~z for market purpose)
Applicant: MEAN WELL Enterprises Co., Ltd.
Address: No. 28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891, Taiwan
Technical Characteristics: NSP-75-xzzzz: AC 100-240V 0.9-0.4A 50/60Hz, AC 277V 0.35A 50/60Hz
NSP-100-xzzzz: AC 100-240V 1.2-0.5A 50/60Hz, AC 277V 0.45A 50/60Hz

The submitted sample of the above equipment has been tested according to following European Directive and standards:

- Electromagnetic Compatibility Directive 2014/30/EU
- MDR Directive (EU) 2017/745

Standards	Report Number	Report date
EN 55032:2015+A1:2020 ; EN 55035:2017+A11:2020 EN IEC 55014-1:2021; EN IEC 55014-2:2021 EN IEC 61000-6-2:2019 EN IEC 61000-3-2:2019+A1:2021+A2:2024 EN 61000-3-3:2013+A1:2019+A2:2021	CE2507WDG0201-1	Sep. 25, 2025
EN 55011:2016+A2:2021 EN 60601-1-2:2015+A1:2021* EN IEC 61000-3-2:2019+A1:2021+A2:2024 EN 61000-3-3:2013+A1:2019+A2:2021	CE2507WDG0201-2	

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the specified European Directive.

This verification does not imply assessment of the production of the product.

*The standard EN 60601-1-2:2015+A1:2021 is under the MDR Directive (EU) 2017/745

Assistant Manager
EMC Department

Name: Madison Luo
Date: Sep. 25, 2025

This document shall not be reproduced, except in full, without the written approval of
Bureau Veritas Shenzhen Co., Ltd. Dongguan Branch.

Information given in this document is related to the tested specimen of the described electrical sample.