



**BUREAU  
VERITAS**

## ATTESTATION of conformity with European Directives

Attestation Number: **2688AB020039001**  
Product: **switching power supplies**  
Brand Name: **MEAN WELL**  
Model: **NSP-1000-xzzzzz (x=12, 15, 24, 27, 36, 48, 60; z=may be blank, -, 0~9, A~Z, a~z for market purpose)**  
Additional Models: **N/A**  
Applicant: **MEAN WELL Enterprises Co., Ltd.**  
Address: **No. 28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891, Taiwan**  
Technical Characteristics: **AC 100-240V 12-5.2A 50/60Hz**  
**AC 277V 4.3A 50/60Hz**

The submitted sample of the above equipment has been tested according to following European Directive and standards:

- Electromagnetic Compatibility Directive 2014/30/EU
- MDR Directive (EU) 2017/745

Standards	Report Number	Report date
EN 55032:2015+A1:2020+A11:2020 EN 55035:2017+A11:2020 EN IEC 55014-1:2021 EN IEC 55014-2:2021 EN 55011:2016+A2:2021 EN 60601-1-2:2015+A1:2021* EN IEC 61000-6-2:2019 EN IEC 61000-3-2:2019+A1:2021+A2:2024 EN 61000-3-3:2013+A1:2019+A2:2021	CE2602WDG0039	Apr. 30, 2026

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the specified European Directive.

This verification does not imply assessment of the production of the product.

\*The standard EN 60601-1-2:2015+A1:2021 is under the MDR Directive (EU) 2017/745

**Assistant Manager  
EMC Department**

**Name: Madison Luo  
Date: Apr. 30, 2026**

This document shall not be reproduced, except in full, without the written approval of  
Bureau Veritas Shenzhen Co., Ltd. Dongguan Branch.

Information given in this document is related to the tested specimen of the described electrical sample.