



## Declaration of Conformity

For the following equipment :

Product Name: Switching Power Supply

Model Designation:RPS-120-X-Y (X=12, 15, 24, 27, 48; Y=C or Blank)

is herewith confirmed to comply with the requirements set out in the Council Directive 93/42/EEC concerning Medical devices, the following standards were applied :

### RoHS Directive (2011/65/EU), (EU)2015/863

### MDR Directive (EU) 2017/745

EN60601-1:2006+A1+A12+A2

TUV certificate No : TA50336473 (RPS-120-x)

EN60601-1:2006+A1+A12+A2

TUV certificate No : TA50335081 (RPS-120-x-C)

EN 60601-1-2:2015+A1:2021

### EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

EN55011:2016+A11:2020

Class B

Harmonic current

EN IEC61000-3-2:2019+A1:2021

Voltage flicker

EN61000-3-3:2013+A1:2019+A2:2021

### EMS (Electro-Magnetic Susceptibility)

ESD air

EN61000-4-2:2009

Level 4

15KV

ESD contact

EN61000-4-2:2009

Level 4

8KV

RF field susceptibility

ENIEC 61000-4-3:2020

Level 3

10V/m(80MHz-2.7GHz)

RF field susceptibility

ENIEC 61000-4-3:2020

Table 9

9~28V/m (385MHz~5.78GHz)

EFT bursts

EN61000-4-4:2012

Level 3

2KV/100KHz

Surge susceptibility

EN61000-4-5:2014+A1:2017

Level 4

2KV/Line-Line

Surge susceptibility

EN61000-4-5:2014+A1:2017

Level 4

4KV/Line-Earth

Conducted susceptibility

EN61000-4-6:2014

Level 3

10V

Magnetic field immunity

EN61000-4-8:2010

Level 4

30A/m

Voltage dip, interruption

ENIEC61000-4-11:2020

0% residual voltage for 0.5 cycles, 0% residual voltage for 1 cycles, 70% residual voltage for 25 cycles, 0% residual voltage for 250 cycles

### Note:

A component power supply with load will be installed into final equipment which consists of an electronically shielded metal enclosure. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure.

For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on <http://www.meanwell.com>)" and TDF (Technical Documentation File).

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC3xxxxxxx

Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/Director, Group R&D :

*Aries*

(Signature)

Alex Tsai/Director, Product Strategy Center :

*[Signature]*

(Name / Position)

(Signature)

Taiwan

Aug. 28th, 2023

(Place)

(Date)